

**Table A5E.** Quality-control summary for matrix-spike recoveries of *N*-Nitrosodimethylamine (NDMA) and 1,2,3-Trichloropropane (1,2,3-TCP) in samples collected for the Southern Sierra Groundwater Ambient Monitoring and Assessment (GAMA) study, California, June 2006.

[Acceptable recovery range is between 70 and 130 percent]

<b>Constituent</b>	<b>Number of spike samples</b>	<b>Minimum recovery (percent)</b>	<b>Maximum recovery (percent)</b>	<b>Median recovery (percent)</b>
1,2,3-Trichloropropane <sup>1</sup>	3	95	114	113
<i>N</i> -Nitrosodimethylamine (NDMA)	3	101	106	104

<sup>1</sup>Constituent detected in ground-water samples