




LINEAR REGRESSION EQUATION

Ca ICP-AES, in mg/L = 1.193 + 1 * Ca cICP-MS, in mg/L; $R^2 = .989$

Confidence Intervals

Ca ICP-AES, in mg/L vs. Ca cICP-MS, in mg/L

	Coefficient	95% Lower	95% Upper
Intercept	1.193	-1.097	3.483
Ca cICP-MS, in mg/L	1.000	.954	1.046