




LINEAR REGRESSION EQUATION

$$\text{Cd ICP-MS, in } \mu\text{g/L} = .002 + .926 * \text{Cd cICP-MS, in } \mu\text{g/L}; R^2 = 1$$

Confidence Intervals Cd ICP-MS, in µg/L vs. Cd cICP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	.002	-.020	.023
Cd cICP-MS, in µg/L	.926	.924	.928