



LINEAR REGRESSION EQUATION

$\text{Cd ICP-MS, in } \mu\text{g/L} = -.002 + .925 * \text{Cd cICP-MS, in } \mu\text{g/L}; R^2 = .997$

Confidence Intervals
Cd ICP-MS, in µg/L vs. Cd cICP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	-.002	-.036	.032
Cd cICP-MS, in µg/L	.925	.904	.946