




LINEAR REGRESSION EQUATION

$$\text{Cr GF-AAS, in } \mu\text{g/L} = -.003 + 1.05 * \text{Cr cICP-MS, in } \mu\text{g/L}; R^2 = .998$$

Confidence Intervals Cr GF-AAS, in µg/L vs. Cr cICP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	-.003	-.489	.482
Cr cICP-MS, in µg/L	1.050	1.029	1.071