




LINEAR REGRESSION EQUATION

$$\text{Cu ICP-MS, in } \mu\text{g/L} = .299 + .964 * \text{Cu clCP-MS, in } \mu\text{g/L}; R^2 = .999$$

Confidence Intervals Cu ICP-MS, in µg/L vs. Cu clCP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	.299	.098	.499
Cu clCP-MS, in µg/L	.964	.953	.975