




LINEAR REGRESSION EQUATION

Fe ICP-AES, in µg/L = -26.485 + 1.078 * Fe clCP-MS, in µg/L; $R^2 = 1$

Confidence Intervals

Fe ICP-AES, in µg/L vs. Fe clCP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	-26.485	-44.006	-8.963
Fe clCP-MS, in µg/L	1.078	1.069	1.087