




LINEAR REGRESSION EQUATION

$$\text{Mg ICP-AES, in mg/L} = -.594 + .993 * \text{Mg cICP-MS, in mg/L}; R^2 = .998$$

Confidence Intervals
Mg ICP-AES, in mg/L vs. Mg cICP-MS, in mg/L

	Coefficient	95% Lower	95% Upper
Intercept	-.594	-1.206	.019
Mg cICP-MS, in mg/L	.993	.975	1.011