




LINEAR REGRESSION EQUATION

$$\text{Mn ICP-MS, in } \mu\text{g/L} = 1.682 + .938 * \text{Mn clCP-MS, in } \mu\text{g/L}; R^2 = .994$$

Confidence Intervals Mn ICP-MS, in µg/L vs. Mn clCP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	1.682	-.026	3.390
Mn clCP-MS, in µg/L	.938	.907	.969