




LINEAR REGRESSION EQUATION

Mo ICP-MS, in µg/L = $-.112 + 1.045 * \text{Mo cICP-MS, in } \mu\text{g/L}$; $R^2 = 1$

Confidence Intervals Mo ICP-MS, in µg/L vs. Mo cICP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	-.112	-.239	.016
Mo cICP-MS, in µg/L	1.045	1.038	1.053