



LINEAR REGRESSION EQUATION

Pb ICP-MS, in µg/L = .024 + .922 * Pb cICP-MS, in µg/L; R² = 1

**Confidence Intervals
Pb ICP-MS, in µg/L vs. Pb cICP-MS, in µg/L**

	Coefficient	95% Lower	95% Upper
Intercept	.024	-.047	.095
Pb cICP-MS, in µg/L	.922	.917	.928