




LINEAR REGRESSION EQUATION

$$\text{Se ICP-MS, in } \mu\text{g/L} = .255 + .831 * \text{Se cICP-MS, in } \mu\text{g/L}; R^2 = .875$$

Confidence Intervals Se ICP-MS, in µg/L vs. Se cICP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	.255	.133	.378
Se cICP-MS, in µg/L	.831	.695	.967