




LINEAR REGRESSION EQUATION

$$\text{Sr ICP-MS, in } \mu\text{g/L} = 68.799 + .87 * \text{Sr cICP-MS, in } \mu\text{g/L}; R^2 = .988$$

Confidence Intervals
Sr ICP-MS, in µg/L vs. Sr cICP-MS, in µg/L

	Coefficient	95% Lower	95% Upper
Intercept	68.799	9.040	128.559
Sr cICP-MS, in µg/L	.870	.829	.911