



LINEAR REGRESSION EQUATION

Sr ICP-MS, in µg/L = -2.933 + 1.004 * Sr cICP-MS, in µg/L; R² = .999

Confidence Intervals
Sr ICP-MS, in µg/L vs. Sr cICP-MS, in µg/L

| | Coefficient | 95% Lower | 95% Upper |
|---------------------|-------------|-----------|-----------|
| Intercept | -2.933 | -6.698 | .831 |
| Sr cICP-MS, in µg/L | 1.004 | .987 | 1.020 |